



ZAKŁAD JAKOŚCI ŻYWNOSCI

92-202 Łódź, Al. Marszałka J. Piłsudskiego 84  
tel. (+48 42) 636 92 11, (+48 42) 636 55 72, (+48 42) 674 64 14 wew. 320, fax (+48 42) 674 81 24  
zj@ibprs.pl  
NIP: 525-000-82-64 REGON: 000053835-00026

Instytut Biotechnologii Przemysłu Rolno-Spożywczego  
Klienci: Prof. Wacława Dąbrowskiego  
02 - 532 Warszawa, ul. Rakowiecka 36  
NIP 525-000-82-64 REGON 000053835  
ZAKŁAD JAKOŚCI ŻYWNOSCI  
92 - 202 Łódź, Al. Marszałka J. Piłsudskiego 84  
tel. (42) 674 64 14, (42) 636 92 11, tel./fax. (42) 674 81 24

1/1

Łódź, 19-08-2020

Certificate of Analysis No K/308/01/2020

**Subject of analysis:** UV-C STERILON FLOW 72W with UVC17H1 lighting (2x36W)

**Klient:** Lena Lighting S.A  
63-000 Środa Wlkp.  
ul. Kórnicka 52

The device for testing was delivered by the Customer: 07-08-2020  
The tests began: 11-08-2020  
The tests finished: 17-08-2020

Type of analysis	Method	Results	
<b>Microbial parameters</b>			
Testing of the level of air pollution during the operation of the lamp in a room of 25 m <sup>2</sup>	Own methodology using a microbiological air sampler MAS-100 ECO™ Instruction MAS-100 Eco™	*[cfu/1 m <sup>3</sup> ]	Reduction of microorganisms
- total viable count of microorganisms at time 0		884	-
- total viable count of microorganisms after 2 hours		225	R <sub>2h</sub> = 71,15%
- total viable count of microorganism after 6 hours		191	R <sub>6h</sub> = 78,45 %
- total viable count microorganisms after 20 hours		21	R <sub>20h</sub> = 97,62%
- number of yeasts and molds and yeasts at time 0		379	-
- number of yeasts and molds and yeasts after 2 hours		101	R <sub>2h</sub> = 73,32%
- number of yeasts and molds and yeasts after 6 hours		76	R <sub>6h</sub> = 80,05 %
- number of yeasts and molds and yeasts after 20 hours		9	R <sub>20h</sub> = 97,75 %

\*\* The results are the average number of microorganisms from two measurements

Authorized:

*B. Paziak-Domańska*  
PRACOWNIA MIKROBIOLOGII  
dr Beata Paziak-Domańska  
Adiunkt

Accepted:

KIEROWNIK ZAKŁADU  
JAKOŚCI ŻYWNOSCI  
*dr Beata Berdziejewska*



**ZAKŁAD JAKOŚCI ŻYWNOSCI**

92-202 Łódź, Al. Marszałka J. Piłsudskiego 84

tel. (+48 42) 636 92 11, (+48 42) 636 55 72, (+48 42) 674 64 14 wew. 320, fax (+48 42) 674 81 24

zj@ibprs.pl

NIP: 525-000-82-64 REGON: 000053835-00026

Instytut Biotechnologii Przemysłu Rolno-Spożywczego  
im. prof. Waława Dąbrowskiego

02 - 532 Warszawa, ul. Rakowiecka 36

NIP 525-000-82-64 REGON 000053835

ZAKŁAD JAKOŚCI ŻYWNOSCI

92 - 202 Łódź, Al. Marszałka J. Piłsudskiego 84

tel. (+48 42) 674 64 14, (+48 42) 636 92 11, fax (+48 42) 674 81 24

**UVC17H1 lighting (2x36W)**

**Assessment of antimicrobial efficacy of air disinfection with UV-C STERILON FLOW 72W with**

### **Aim and scope of the study**

The aim of the study was to determine the effectiveness of air disinfection with UV-C STERILON FLOW 72W with UVC17H1 lighting (2x36W)

(Certificate of Analysis K/308/01/2020) on the basis of the total viable count of microorganisms and number of molds and yeasts examination using aspiration method after 2, 6 and 20 hours flow UVC lamp working in a room with an area of 25 m<sup>2</sup>.

### **Experimental procedure**

The studies were conducted in accordance with its own methodology developed at the Laboratory and the manufacturer's manual MAS-100 ECO<sup>TM</sup> (Microbiological Air Sampler) in a room with an area of 25 m<sup>2</sup>. Before turning on the lamp, the total viable count of microorganisms and the number of mold and yeast in the room air were examined (at 0 time). The flow UVC lamp was placed in the center of the room and the air pollution was measured 2 meters from the device after 2, 6 and 20 hours of operation. The tests were carried out using the aspiration method using the microbiological air sampler MAS-100 ECO<sup>TM</sup>. Each time the device was placed on a flat surface, at a height of about 90 cm from the floor, and took 1000 liters of air through a perforated plate. The air stream containing particles was directed to the PCA or YGC agar surface in a standard Petri dish. After completing the air sampling cycle, the Petri dishes were incubated at 30°C for 72h or 25°C for 5 days, then the colonies grown were counted and the number of microorganisms in 1 m<sup>3</sup> of air was determined, taking into account the correction of the Feller's statistical correction table.

*B. Paziak-Domańska*

PRACOWNIA MIKROBIOLOGII  
dr Beata Paziak-Domańska  
Adiunkt